





REVISION HISTORY				
ZONE	REV	DESCRIPTION	DATE	APPROVED
FAN LEVEL	1	UL FAN MODULE	7/06/2010	P CH
MAIN FILTER LEVEL	2	METAL CLAMPS	8/12/2011	P CH
RIGHT SIDE	3	METAL SHOEBOX	3/28/2012	P CH
SIDE GLASSES	4	AIR METER HOLES (2)	1/5/2017	P CH
TOP COVER	5	UL FAN METAL MODULE	10/31/2017	P CH



- NOTES:**
- 1.0 CUSTOMER:
 - 2.0 SUBMIT DRAWINGS FOR APPROVAL.
 - 3.0 INCLUDED;
 - 3.1 ENCLOSURE BUILT WITH 3/8" WHITE POLIPROPYLENE, SEAMLESS, NON-POROUS WATERTIGHT INSIDE, THERMALLY FUSE.
 - 3.2 FRONT SASH, SIDES PANELS MADE IN CLEAR ACRYLIC OF 1/4".
 - 3.3 SWITCHES MAIN ON/OFF.
 - 3.4 MONITORING LOW AIRFLOW
 - 4.0 AIRMETER (OPTIONAL).
 - 5.0 SAFETY FILTER (OPTIONAL).
 - 6.0 INTERNAL MANUAL SPEED CONTROLLER.
 - 7.0 INTERNAL LIGHTNING.
 - 8.0 ELECTROSTATIC PRE-FILTER.
 - 9.0 STANDARD MOLDED PLUG NEMA 5-15P
 - 10.0 ELECTRICAL REQUIREMENTS 110V 60HZ.

A-A (1 : 8)



B-B (1 : 8)

Designed by CFG	Checked by	Approved by	Date 8/26/2015	Date
AIR SCIENCE USA			PURAIR P10BT-PP	
PP10BTUL.GAssem			Edition 5	Sheet 1 / 1