

CERTIFICATE OF TEST & CALIBRATION

Appliance	Ki-Discus Mk. 2		Serial Number		
Customer					
Date of Issue			Certificate Number	CTQ	
Temperature	°C	Humidity	%	Supply Voltage	v
Test Engineer					
<p>The unit under test will be subject to the Standard Ki-Discus test procedures. These procedures cover (1) Nebuliser Nozzle/Disc Clearance (2) Disc Rotation Speed (3) Mechanical Gauge Reading (4) Pump Delivery.</p>					
TEST	PARAMETERS		SET RESULTS		
(1) Nebuliser/Disc Clearance	0.10 mm ± 0.05mm				
(2) Disc Rotation Speed	28,000 RPM ± 250 RPM				
(3) Vacuum Gauge	-20.0 mbar ± 0.5 mbar				
(4) Pump Delivery	20.0 ml ± 0.25ml in 9m 00s ±15s				
<p>This equipment has passed all Standard Internal Functional and Calibration Tests and is compliant with the product specification.</p>					
DATE OF CALIBRATION					
NEXT CALIBRATION DUE DATE					



Approved Signatory _____

Signature *RAult* _____

Test Equipment	Calibration Company	Serial no.	Calibration Expiry Date
(1) External Micrometer	RS Components	161183527	07 August 2025
(1) Feeler Gauge	RS Components	1931787/252	08 August 2025
(2) Frequency Counter	Compli Group Ltd.	498974	30 July 2025
(3) Digital Manometer	Labcal Ltd	57235	01 April 2026

CTS (Europe) Ltd certifies the above instruments meet correct specifications and have been calibrated using instruments of known accuracies, which are traceable to UKAS accreditation.

Uncertainties

Reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor K=2, providing a level of confidence of approximately 95%. The Uncertainty evaluation has been carried out in accordance to UKAS Requirements - Digital Manometer: Uncertainty is [0.01% of the applied pressure + 0.43 mbar (+ resolution of the instrument of 0.1 mbar)]. The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. Frequency counter: Measurement uncertainties in range: 10 Hz to 100 kHz: ± (1 in 10⁶ + 1 L.S.D.). Uncertainties quoted refer to values applied under test. The L.S.D. component of the above measurement uncertainties refers to the display resolution of the instrument under test. Feeler gauge: Uncertainty of measurement: ± 0.002 mm. Digital ext. Micrometer: Uncertainty of measurement: Micrometer : ± 0.002 mm, flatness: ± 0.0003 mm, parallelism: ±0.0008 mm.